


For ToxiLight™ Non-destructive Cytotoxicity Assay Lucetta™ Luminometer Quick Start Guide


Supporting operating instructions for the Lucetta™ Luminometer are available on the accompanying CD or at www.lonza.com/lucetta. Full kit instructions can be downloaded from www.lonza.com/toxilight.

Working Step

1. Remove cells from incubator and allow cooling down to room temperature.
2. Reconstitute AK Detection Reagent (AKDR) in Assay Buffer Equilibrate **15 minutes** at room temperature
3. Transfer 20 µl of suspension cells or media from adherent cells into a fresh luminometer tube.
4. Select Program "RUN Single Read" and press ENTER.




5. Add 100 µl AKDR to the sample tube. Wait **5 minutes**.
6. Place tube in holder. Close Door.



Working Step

7. Luminescence measurement automatically starts after a delay of 2 s.


The default measuring time is 1 s



8. See measured RLU/s value for sample 1.

Press CONTINUE to move to next sample.

Press ESC to return to main screen.



Contact Information

North America

Customer Service: 800 638 8174 (toll free)
order.us@lonza.com
Scientific Support: 800 521 0390 (toll free)
scientific.support@lonza.com

Europe

Customer Service: +32 87 321 611
order.europe@lonza.com
Scientific Support: +32 87 321 611
scientific.support.eu@lonza.com

International

Contact your local Lonza distributor
Customer Service: +1 301 898 7025
scientific.support@lonza.com

International Offices

Australia	+61 3 9550 0883
Belgium	+32 87 321 611
Brazil	+55 11 2069 8800
France	0800 91 19 81 (toll free)
Germany	0800 182 52 87 (toll free)
India	+91 40 4123 4000
Ireland	1 800 654 253 (toll free)
Italy	800 789 888 (toll free)
Japan	+81 3 6264 0660
Luxemburg	+32 87 321 611
Poland	+48 781 120 300
Singapore	+65 6521 4379
Spain	900 963 298 (toll free)
The Netherlands	0800 022 4525 (toll free)
United Kingdom	0808 234 97 88 (toll free)

Lonza Cologne GmbH – 50829 Cologne Germany

For research use only. Not for use in diagnostic procedures. Unless otherwise noted, all trademarks herein are marks of the Lonza Group or its affiliates. The information contained herein is believed to be correct and corresponds to the latest state of scientific and technical knowledge. However, no warranty is made, either expressed or implied, regarding its accuracy or the results to be obtained from the use of such information and no warranty is expressed or implied concerning the use of these products. The buyer assumes all risks of use and/or handling. No statement is intended or should be construed as a recommendation to infringe any existing patent.

© Copyright 2012, Lonza Cologne GmbH. All rights reserved.
FL-QSGToxilight 06/12 CD-MN011

